



CPC, Taiwan - Ichthys, Australia Correlation Test 2025



COMPARISON RESULTS BETWEEN TERMINALS

Company	Terminal	GC No.	O2	N2	CH4	C2H6	CO2	C3H8	i-C4H10	n-C4H10	i-C5H12	n-C5H12	C6H14
CPC	Yung-An	H											
		L											
INPEX	Ichthys Lab	08	0.01	0.39	84.77	13.71	0.01	0.97	0.05	0.02	0.05	0.02	0.00
		09	0.01	0.39	84.74	13.72	0.01	0.99	0.05	0.02	0.05	0.02	0.00
	Ichthys Lab	L410-AT-1001	0.01	0.40	84.70	13.75	0.01	0.99	0.05	0.02	0.05	0.02	0.00
		L410-AT-0001	0.01	0.39	84.72	13.74	0.01	0.99	0.05	0.02	0.05	0.02	0.00
Max. - Min.			0.00	0.01	0.07	0.04	0.00	0.02	0.00	0.00	0.00	0.00	0.00
Tolerance			0.00	0.03	0.59	0.27	0.00	0.02	0.00	0.00	0.00	0.00	0.00